



This drawing and the information contained thereon remain the property of BeaconMedaes and may not be used for other than the purpose for which it is loaned without the expressed written permission from BeaconMedaes Engineering.

Parts List			
ITEM	QTY	PART NUMBER	DESCRIPTION
1	1	4107 7847 34	ENCLOSURE ASM CLAW VSD PX
1.1	1	4107 6512 50	CAB CSD603610 60x36x10 HX
1.2	2	4107 6509 94	PLUG AS050 HOLE SEAL 1/2" CONDUIT
1.3	1	4107 6509 99	PLUG AS200 HOLE SEAL 2" CONDUIT
1.4	3	4107 6507 81	HANDLE HM1R12 ROTARY THRU-DOOR
1.7	1	4107 6516 75	DECAL OVERLAY DOOR DISPLAY STD PX
1.8	1	4107 6523 84	DECAL OVERLAY DISCONNECT PX VSD
2	1	4107 7844 54	TEMP PROBE ASM PNL AMBIENT
3.5	1	4107 6516 60	PCB PIR MOTION SENSOR
3.8	1	4107 6507 35	LED POWER GREEN 24VDC
3.9	1	4107 6508 38	HORN AUDIBLE ALARM 24VDC

REFERENCE	DETAIL "A" DESCRIPTION
RTD-7	PANEL AMBIENT TEMPERATURE PROBE
PUMP1	
RTD-1	HI-DISCHARGE TEMPERATURE 1
FN-1	PUMP FAN 1 (15HP ONLY)
PUMP 2	
RTD-2	HI-DISCHARGE TEMPERATURE 2
FN-2	PUMP FAN 2 (15HP ONLY)
PUMP 3	
RTD-3	HI-DISCHARGE TEMPERATURE 3
FN-3	PUMP FAN 3 (15HP ONLY)
PUMP 4	
RTD-4	HI-DISCHARGE TEMPERATURE 4
FN-4	PUMP FAN 4 (15HP ONLY)
PUMP 5	
RTD-5	HI-DISCHARGE TEMPERATURE 5
FN-5	PUMP FAN 5 (15HP ONLY)
RECEIVER	
TDC-1	VACUUM TRANSDUCER
BVS-1	BACKUP VACUUM SWITCH





This drawing and the information contained thereon remain the property of BeaconMedas and may not be used for other than the purpose for which it is loaned without the expressed written permission from BeaconMedas Engineering.

Parts List				
ITEM	QTY	PART NUMBER	DESCRIPTION	
3	1	4107 7848 24	PANEL DISPLAY ASM STD PX	
3.1	1	4107 6516 33	GASKET PANEL DOOR DISPLAY STD PX	
3.2	9	0301 2318 00	FLAT WASHER 5mm	
3.3	9	0333 3100 04	WASHER EXT. TOOTH M5	
3.4	9	5537 3448 00	HEX NUT 10-32	
3.5	1	4107 6516 60	PCB PIR MOTION SENSOR	
3.6	1	4107 6506 57	CLAMP PCB PIR MOTION SENSOR	
3.7	2	4107 6516 98	HDW NUT M5 NYLON SPEED	
3.8	1	4107 6507 35	LED POWER GREEN 24VDC	
3.9	1	4107 6508 38	HORN AUDIBLE ALARM 24VDC	
3.10	1	4107 7847 88	DISPLAY LCD ASM 5.7	
3.11	5	4107 7847 84	DISPLAY LCD ASM 3.5	
3.12	1	4107 6519 26	CABLE PCB TO PIR SENSOR	
3.13	1	4107 6519 23	CABLE PCB TO HORN	
3.14	8	4107 6516 95	HDW M-M HEX STANDOFF 10-32 x 3/4"L	
3.15	1	4107 6516 52	COVER PANEL DOOR INT. STD PX	
4.19	11	4107 6516 98	HDW NUT M5 NYLON SPEED	
4.20	1	4107 6508 94	WIRE DUCT SLHD1.5X2G4 1.5"x2"x11.25"L	
4.21	1	4107 6509 01	COVER TC1.5-G 1.5"Wx11.25"L DUCT	
4.22	1	4107 6508 94	WIRE DUCT SLHD1.5X2G4 1.5"Wx2"x31.18"L	
4.23	1	4107 6509 01	COVER TC1.5-G 1.5"Wx31.18"L DUCT	
5	1	4107 7844 56	TEMP PROBE ASM CABINET INTERNAL	

DETAIL A
GASKET, PIR MOTION SENSOR,
& ALARM HORN ASSEMBLY

DETAIL C
3.5 LCD DISPLAY/
TOUCHSCREEN ASSEMBLY
(5 REQUIRED FOR PENTAPLEX)



DETAIL B
5.7 LCD DISPLAY
TOUCHSCREEN ASSEMBLY

Revised:	Date:	Drawn:	Date:
		JAH	11/22/2010
INSPECTION CHECKLIST UNLESS OTHERWISE SPECIFIED:		UNLESS OTHERWISE SPECIFIED ALL DIMENSIONS ARE IN INCHES AND TOLERANCES ARE: XXX ±.02 XXX±.005	
ALL ITEMS MARKED BY ONE OF THE ABOVE SYMBOLS ARE TO BE INSPECTED		Description: DWG PANEL CLAW VSD PX 15HP DUAL ALARM CONTACTS INTERNAL DOOR	
EQUIPMENT TO BE USED DURING RECEIVING INSPECTION SHALL BE PER QUALITY SYSTEM PROCEDURES		Part Number: 4107 8517 85	
		DN: HOP 100749 01	Rev: 00
		Sheet 2 of 3	DO NOT SCALE THIS DOCUMENT


This drawing and the information contained thereon remain the property of BeaconMedaes and may not be used for other than the purpose for which it is loaned without the expressed written permission from BeaconMedaes Engineering.

Notes:


1. COMPONENT LABELS TO BE PLACED ON BACKPLATE ONLY.



DETAIL A



DETAIL B



DETAIL C

Parts List				
ITEM	QTY	PART NUMBER	DESCRIPTION	
4	1	4107 7832 25	WIRE DUCT ASM CLAW VSD PX	
6	1	SEE BOM	DIST BLOCK ASM 3P	
7	5	SEE BOM	POWER SUPPLY ASM 24V 60W	
8	1	4107 7856 57	FUSE BLOCK ASM 5x20 CLAW VSD PX	
9	1	4107 7845 02	RELAY ASM AIRVAC STD PX	
10	1	4107 7862 78	TERMINAL BLOCK ASM CLAW VSD PX 2 AUX	
11	1	4107 7862 77	TERMINAL BLOCK ASM CLAW VSD PX FLD	
12	1	4107 7847 92	MASTER PCB ASM FULL I/O	
13	3	4107 7847 49	ETHERNET SECURITY PLUG ASM	
14	1	4107 7848 95	PUMP PCB ASM 16 I/O PX	
15	2	SEE BOM	DISCONNECT ASM FD 3-POLE	
16	6	SEE BOM	CIRCUIT BREAKER ASM EHD 3-POLE	
17	1	SEE BOM	VSD ASM 15HP 208V	
18	5	SEE BOM	STARTER/OVERLOAD ASM FVR	

INSPECTION CHECKLIST
UNLESS OTHERWISE SPECIFIED:
ALL ITEMS MARKED BY ONE OF THE ABOVE SYMBOLS ARE TO BE INSPECTED
EQUIPMENT TO BE USED DURING RECEIVING INSPECTION SHALL BE PER QUALITY SYSTEM PROCEDURES

Revised:	Date:	Drawn: JAH	Date: 11/22/2010
UNLESS OTHERWISE SPECIFIED ALL DIMENSIONS ARE IN INCHES AND TOLERANCES ARE: XXX ±.02 XXX ±.005		Scale: NTS	
Description: DWG PANEL CLAW VSD PX 15HP DUAL ALARM CONTACTS BACKPLATE		Part Number: 4107 8517 85	
DN: HOP 100749 01		Rev: 00	
Sheet 3 of 3		DO NOT SCALE THIS DOCUMENT	